



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/11/2006 INITIALS: DEC OWNER: WSP  
DISTRICT: 7 ADDRESS: \_\_\_\_\_  
LOCATION: 7 M Shop COUNTY: \_\_\_\_\_  
MAKE: Haenni TYPE: port wheel weigher ACCURACY CLASS: III  
MODEL: 101 ID. NO: 34143 CAPACITY: 20000 SECTIONS: 1  
MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15000	0				
North	20000	20050	+50				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cowan

Date: 4/18/2006

State of Washington

County of Thurston


Signed and sworn to (or affirmed) before  
me on 4/18/2006 by

Michael J. Evans  
(Signature of notary public)

Michael J. Evans

Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS